

地址 (Add): 湖北省武汉市东湖新技术开发区茅店山中路二号  
邮编 (Post Code): 430223  
电话 (Tel): 027-81925136  
网址 (Web site): http://www.hbjl.gov.cn  
传真 (Fax): 027-81925137

CUC

REGISTRATION NO. 002


# 校准证书

## CALIBRATION CERTIFICATE

证书编号: [2017DW00860427]  
Certificate No.

委托方名称 Customer	中国水电建设集团十五工程局有限公司
委托方地址 Address	西安市雁塔区科技路16号
器具名称 Name Of Sample	发电机转子交流阻抗测试仪
制造厂商 Manufacturer	武汉世纪华胜科技有限公司
型号规格 Model/Type	FS-30ZK
器具编号 No Of Sample	306170304

证书专用章  
Stamp



校准日期 Calibration date	2017 年 05 月 16 日 Y M D
建议再校日期 Recommended recal.date	2018 年 05 月 15 日 Y M D

证书批准人  
Approved by 耿靖

核验员  
Checked by 李艳

校准员  
Calibrated by 尹伟

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



● 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

● 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

● 本次校准的技术依据 (名称、代号)

Reference documents for the Calibration (Name、Code)

参照: JJF(鄂) 11—2010 数字多用表校准规范

● 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称  
Name of Equipment  
交直流标准表

型号/编号  
Model/Serial No.  
9080A/015

证书号/有效期  
Certificate No./Due Date  
2016DW00860398/2017-05-31

● 校准环境条件

Environmental condition on the Calibration

温度: 20.9℃

Temperature

气压: —

Pressure

相对湿度: 48 %

R.H.

地点: 本院光谷基地B102

Place

其它: —

Others

原始记录编号: 2017DW00860427

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.

未经本院许可, 不得部分复制本证书。

量测试技  
书骑缝章

### 校准数据/结果

Data/Results of Calibration

被校显示值 (V)	标准值 (V)	被校显示值 (A)	标准值 (A)
50.75	50.65	1.003	1.009
100.2	100.2	2.015	2.017
200.8	200.9	5.03	5.03
300.5	300.7	10.18	10.19
400.4	400.4	20.04	20.04
---	---	30.10	30.13
测量不确定度	电压: $U_{rel}=0.1\%$ ; $k=2$		电流: $U_{rel}=0.1\%$ ; $k=2$

以下空白

研究院